

EX-TROL™ Plus  

Hematology Control

LOT XP038

 **2008-05-02**

Quality Control Months: March, April

DREW EXCELL 22	LOT XP038L				LOT XP038N				LOT XP038H			
	Mean	Limits	Range		Mean	Limits	Range		Mean	Limits	Range	
WBC K/uL	3.2	± 0.5	2.7	3.7	8.1	± 0.8	7.3	8.9	23.2	± 2.5	20.7	25.7
Neut K/uL	1.6	± 0.4	1.2	2.0	4.9	± 0.6	4.3	5.5	17.4	± 1.5	15.9	18.9
Neut %	51.5	± 8.0	43.5	59.5	61.0	± 6.0	55.0	67.0	75.0	± 6.0	69.0	81.0
Lymp K/uL	1.1	± 0.4	0.7	1.5	2.1	± 0.6	1.5	2.7	3.1	± 1.3	1.8	4.4
Lymp %	34.0	± 7.0	27.0	41.0	26.0	± 6.0	20.0	32.0	13.5	± 5.0	8.5	18.5
Mono K/uL	0.3	± 0.3	0.0	0.6	0.7	± 0.5	0.2	1.2	1.4	± 1.1	0.3	2.5
Mono %	8.0	± 6.0	2.0	14.0	8.5	± 4.5	4.0	13.0	6.0	± 4.0	2.0	10.0
Eos K/uL	0.2	± 0.2	0.0	0.4	0.3	± 0.3	0.0	0.6	1.0	± 1.0	0.0	2.0
Eos %	5.0	± 5.0	0.0	10.0	3.5	± 3.5	0.0	7.0	4.5	± 4.0	0.5	8.5
Baso K/uL	0.1	± 0.1	0.0	0.2	0.1	± 0.1	0.0	0.2	0.2	± 0.2	0.0	0.4
Baso %	1.5	± 1.5	0.0	3.0	1.0	± 1.0	0.0	2.0	1.0	± 1.0	0.0	2.0
RBC M/uL	2.06	± 0.15	1.91	2.21	4.60	± 0.20	4.40	4.80	5.25	± 0.24	5.01	5.49
HGB g/dL	5.5	± 0.4	5.1	5.9	13.7	± 0.6	13.1	14.3	16.7	± 0.8	15.9	17.5
HCT %	16.9	± 2.0	14.9	18.9	41.4	± 2.4	39.0	43.8	49.9	± 2.8	47.1	52.7
MCV fL	82.0	± 5.0	77.0	87.0	90.0	± 5.0	85.0	95.0	95.0	± 5.0	90.0	100.0
MCH pg	26.7	± 2.4	24.3	29.1	29.8	± 2.8	27.0	32.6	31.8	± 2.8	29.0	34.6
MCHC g/dL	32.6	± 3.0	29.6	35.6	33.1	± 3.0	30.1	36.1	33.5	± 3.0	30.5	36.5
RDW %	13.7	± 4.0	9.7	17.7	13.3	± 4.0	9.3	17.3	12.6	± 4.0	8.6	16.6
PLT K/uL	76	± 20	56	96	233	± 30	203	263	474	± 60	414	534
MPV fL	7.6	± 2.0	5.6	9.6	7.3	± 2.0	5.3	9.3	7.7	± 2.0	5.7	9.7