



DREW Scientific, Inc.

Product No. 610020

Lot No. CDC049

Expiration 5-May-2009

MD CAL-KIT™

HEMATOLOGY CALIBRATOR HEMAVET SERIES

For use with HEMAVET® HEMATOLOGY INSTRUMENTS

Parameter (units)	Target	Range
WBC (K/ μ L)	9.8	9.4 – 10.2
RBC (M/ μ L)	4.68	4.51 – 4.85
Hb (g/dL)	13.4	13.0 – 13.8
MCV (fL)	80	78.5 – 81.5
PLT (K/ μ L)	275	250 – 300

IMPORTANT

Follow mixing instructions from reverse side and
Wipe threads on bottle and cap before closing

**FOR IN VITRO DIAGNOSTIC USE
PROTECT FROM FREEZING
STORE AT 2° - 8° C (35° - 46° F)**

POTENTIAL BIOHAZARDOUS MATERIAL

4230 Shilling Way Dallas, Texas. 75237 Phone (214) 210-4900
Fax: (214) 210-4949
Web Site: drew-scientific.com