DREW Scientific, Inc.
2056 THOMASTON AVENUE
Waterbury, CT 06704 ● USA

DREW Scientific, Inc.

MULTI -TROL ™

Multispecies Hematology Reference Control

Species Key: CAT Product Number: 600064, 600069

Lot Number: 1624 Expiration Date: 04-15-08

For use with HEMAVET ® **and** MASCOT ™ **Hematology Analyzers**

IMPORTANT

Follow mixing instructions from reverse side and wipe threads on bottle and cap before closing

FOR IN VITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2° - 8° C (35° - 46° F)

	HEMAVET 850, 950 MASCOT 850		HEMAVET 550 MASCOT 550		HEMAVET 950 REV F and Above		HEMAVET 800 MASCOT 800		HEMAVET 500/300		HEMAVET 600 MASCOT V		HEMAVET 400, 200		HEMAVET 1500, 1700, 2500, 3500, 3700	
Parameter (Units)	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**	**Target**	**Range**
WBC (K/µL)	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5	7.9	7.3 – 8.5
NE (K/µL)																
NE/MO/BA (K/µL)																
GR/MO (K/µL)																
LY (K/µL)	1.6	1.2 – 2.2	1.6	1.2 – 2.2	1.6	1.2 – 2.2	1.6	1.2 – 2.2	1.6	1.2 – 2.2	0.9	0.5 – 1.4	0.9	0.5 – 1.4	1.6	1.2 – 2.2
MO (K/µL)																
EO (K/µL)																
BA (K/µL)																
NRBC (K/µL)																
NE (%)																
NE/MO/BA (%)																
GR/MO (%)																
LY (%)	20.8	15.8 – 25.8	20.8	15.8 – 25.8	20.8	15.8 – 25.8	20.8	15.8 – 25.8	20.8	15.8 – 25.8	11.8	6.8 – 16.8	11.8	6.8 – 16.8	20.8	15.8 – 25.8
MO (%)																
EO (%)																
BA (%)																
NRBC (%)																
RBC (M/	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70	7.38	7.06 – 7.70
Hb (g/dL)	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5	12.9	12.3 – 13.5
HCT (%)	35.9	31.6 – 40.6	35.9	31.6 – 40.6	40.9	37.9 – 43.9	35.9	31.6 – 40.6	35.9	31.6 – 40.6	35.9	31.6 – 40.6	35.9	31.6 – 40.6	35.9	31.6 – 40.6
MCV (fL)	48.7	44.7 – 52.7	48.7	44.7 – 52.7	55.4	51.4 – 59.4	48.7	44.7 – 52.7	48.7	44.7 – 52.7	48.7	44.7 – 52.7	48.7	44.7 – 52.7	48.7	44.7 – 52.7
MCH (pg)	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1	17.5	16.0 – 19.1
MCHC (g/dL)	35.9	30.3 – 42.8	35.9	30.3 – 42.8	31.6	29.1 – 34.1	35.9	30.3 – 42.8	35.9	30.3 – 42.8	35.9	30.3 – 42.8	35.9	30.3 – 42.8	35.9	30.3 – 42.8
RDW (%)	22.9	20.9 – 24.9	22.9	20.9 – 24.9	23.3	21.3 – 25.3	22.9	20.9 – 24.9	22.9	20.9 – 24.9	22.9	20.9 – 24.9	22.9	20.9 – 24.9	22.9	20.9 – 24.9
Retics (M/µL)																
Retics (%)																
PLT (K/µL)	182	117 - 247	182	117 - 247	182	117 - 247	182	117 - 247	182	117 - 247	182	117 - 247	182	117 - 247	182	117 - 247