



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: DOG Product Number: 600066, 600068

Lot: 1666 Expiration Date: 4-15-09

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	12.5	11.9 - 13.1	12.5	11.9 - 13.1
NE (K/μL)	9.4	8.0 - 10.9		
NE/MO/BA (K/μL)			10.1	8.7 - 11.7
LY (K/μL)	1.4	0.7 - 2.1	1.4	0.7 - 2.1
MO (K/μL)	0.4	0.0 - 0.7		--Elev
EO (K/μL)	1.0	0.4 - 1.8	1.0	0.4 - 1.8
BA (K/μL)	0.3	0.0 - 0.6		--Elev
NRBC (K/μL)				
NE (%)	75.3	67.3 - 83.3		
NE/MO/BA (%)			80.82	72.8 - 88.8
LY (%)	10.8	5.8 - 15.8	10.8	5.8 - 15.8
MO (%)	2.8	0.5 - 5.1		--Elev
EO (%)	8.4	3.4 - 13.4	8.4	3.4 - 13.4
BA (%)	2.7	0.7 - 4.7		--Elev
NRBC (%)				
RBC (M/μL)	6.01	5.72 - 6.30	6.01	5.72 - 6.30
HB (g/dL)	15.9	15.3 - 16.5	15.9	15.3 - 16.5
HCT (%)	40.5	36.2 - 44.9	40.5	36.2 - 44.9
MCV (fL)	67.3	63.3 - 71.3	67.3	63.3 - 71.3
MCH (pg)	26.4	24.3 - 28.8	26.4	24.3 - 28.8
MCHC (g/dL)	39.3	34.0 - 45.5	39.3	34.0 - 45.5
RDW (%)	17.9	15.9 - 19.9	17.9	15.9 - 19.9
PLT (K/μL)	416	351 - 481	416	350.9 - 480.9

DREW SCIENTIFIC, INC. 4230 Shilling Way Dallas, TX 75237-1023 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2^o - 8^o (35^o - 46^o)