



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: DOG

Product Number: 600066, 600068

Lot: 1676

Expiration Date: 7-15-09

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	12.7	12.1 - 13.3	12.7	12.1 - 13.3
NE (K/μL)	9.5	8.1 - 11.0		
NE/MO/BA (K/μL)			10.0	8.6 - 11.6
LY (K/μL)	1.2	0.5 - 1.9	1.2	0.5 - 1.9
MO (K/μL)	0.3	0.0 - 0.7		--Elev
EO (K/μL)	1.5	0.8 - 2.3	1.5	0.8 - 2.3
BA (K/μL)	0.2	0.0 - 0.5		--Elev
NRBC (K/μL)				
NE (%)	74.6	66.6 - 82.6		
NE/MO/BA (%)			78.90	70.9 - 86.9
LY (%)	9.1	4.1 - 14.1	9.1	4.1 - 14.1
MO (%)	2.6	0.3 - 4.9		--Elev
EO (%)	12.0	7.0 - 17.0	12.0	7.0 - 17.0
BA (%)	1.7	0.0 - 3.7		--Elev
NRBC (%)				
RBC (M/μL)	6.02	5.73 - 6.31	6.02	5.73 - 6.31
HB (g/dL)	15.2	14.6 - 15.8	15.2	14.6 - 15.8
HCT (%)	42.4	38.0 - 46.9	42.4	38.0 - 46.9
MCV (fL)	70.4	66.4 - 74.4	70.4	66.4 - 74.4
MCH (pg)	25.2	23.1 - 27.6	25.2	23.1 - 27.6
MCHC (g/dL)	35.9	31.1 - 41.5	35.9	31.1 - 41.5
RDW (%)	19.2	17.2 - 21.2	19.2	17.2 - 21.2
PLT (K/μL)	416	351 - 481	416	350.7 - 480.7

DREW SCIENTIFIC, INC. 4230 Shilling Way Dallas, TX 75237-1023 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2^o - 8^o (35^o - 46^o)