



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: MOUSE Product Number: 600065, 600070

Lot 1695 Expiration Date: 01-15-2010

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	19.3	17.8 - 20.8	19.3	17.8 - 20.8
NE (K/μL)	6.8	4.8 - 9.0		
NE/MO/BA (K/μL)			10.4	8.4 - 12.7
LY (K/μL)	7.5	6.0 - 9.1	7.5	6.0 - 9.1
MO (K/μL)	0.3	0.0 - 0.8		--Elev
EO (K/μL)	1.4	0.0 - 2.1	1.4	0.0 - 2.1
BA (K/μL)	3.3	2.3 - 4.4		--Elev
NRBC (K/μL)				
NE (%)	35.1	27.1 - 43.1		
NE/MO/BA (%)			54.12	47.1 - 61.1
LY (%)	38.9	33.9 - 43.9	38.9	33.9 - 43.9
MO (%)	1.7	0.0 - 4.0		--Elev
EO (%)	7.2	0.0 - 10.2	7.2	0.0 - 10.2
BA (%)	17.2	13.1 - 21.3		--Elev
NRBC (%)				
RBC (M/μL)	8.68	8.29 - 9.07	8.68	8.29 - 9.07
HB (g/dL)	14.7	14.1 - 15.3	14.7	14.1 - 15.3
HCT (%)	45.6	40.2 - 51.3	45.6	40.2 - 51.3
MCV (fL)	52.5	48.5 - 56.5	52.5	48.5 - 56.5
MCH (pg)	16.9	15.5 - 18.4	16.9	15.5 - 18.4
MCHC (g/dL)	32.2	27.4 - 38.0	32.2	27.4 - 38.0
RDW (%)	23.0	21.0 - 25.0	23.0	21.0 - 25.0
PLT (K/μL)	727	567 - 887	727	567 - 887

DREW SCIENTIFIC, INC. 4230 Shilling Way Dallas, TX 75237-1023 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2° - 8° (35° - 46°)