



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: MOUSE Product Number: 600065, 600070

Lot 1705 Expiration Date: 04-15-2010

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	18.4	16.9 - 19.9	18.4	16.9 - 19.9
NE (K/μL)	9.3	7.2 - 11.7		
NE/MO/BA (K/μL)			10.8	8.7 - 13.1
LY (K/μL)	7.2	5.7 - 8.7	7.2	5.7 - 8.7
MO (K/μL)	0.4	0.0 - 0.9		--Elev
EO (K/μL)	0.5	0.0 - 1.1	0.5	0.0 - 1.1
BA (K/μL)	1.1	0.3 - 2.0		--Elev
NRBC (K/μL)				
NE (%)	50.7	42.7 - 58.7		
NE/MO/BA (%)			58.53	51.5 - 65.5
LY (%)	38.8	33.8 - 43.8	38.8	33.8 - 43.8
MO (%)	2.0	0.0 - 4.3		--Elev
EO (%)	2.6	0.0 - 5.6	2.6	0.0 - 5.6
BA (%)	5.8	1.7 - 9.9		--Elev
NRBC (%)				
RBC (M/μL)	8.56	8.17 - 8.95	8.56	8.17 - 8.95
HB (g/dL)	15.4	14.8 - 16.0	15.4	14.8 - 16.0
HCT (%)	50.4	44.9 - 56.3	50.4	44.9 - 56.3
MCV (fL)	58.9	54.9 - 62.9	58.9	54.9 - 62.9
MCH (pg)	18.0	16.5 - 19.6	18.0	16.5 - 19.6
MCHC (g/dL)	30.5	26.3 - 35.6	30.5	26.3 - 35.6
RDW (%)	23.2	21.2 - 25.2	23.2	21.2 - 25.2
PLT (K/μL)	726	566 - 886	726	566 - 886

DREW SCIENTIFIC, INC. 4230 Shilling Way Dallas, TX 75237-1023 • USA

Revised 13-Jan-10

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2° - 8° (35° - 46°)